



Accredited to LST EN ISO / IEC 17025: 2018
ELGAMA-ELEKTRONIKA CALIBRATION AND VERIFICATION LABORATORY
(Visorių str. 2, LT-08300 Vilnius)

SCOPE OF ACCREDITATION

Calibration and measurement capability (CMC) expressed as:				
Measurand	Reference number of calibration method or procedure	Type of instrument to be calibrated	Measurement range and additional parameters (where applicable)	Measurement uncertainty
Power	KI 02.02-02, 7 edition, 20.11.2017	Operating power standards and test benches for electricity meters	$U = (30 \div 300) \text{ V}$ $f = (45 \div 65) \text{ Hz}$ $\varphi = (0 \div 360)^\circ$ $I = (0.005 \div 120) \text{ A}$	0,03%

Director

Dalia Baležentė

Note. In case of any discrepancies, ambiguities or disputes regarding the subject matter content between the English and Lithuanian versions of the document, the Lithuanian version shall prevail.